
LMG 1106a

Ledwell

Cruise Data Report

By W. Kevin Pedigo & Paul Queior



1 Aug, 2011 – 20 Aug, 2011

Table of Contents

INTRODUCTION	1
ARCHIVE DATA EXTRACTION	2
CD DIRECTORY STRUCTURE.....	3
DISTRIBUTION CONTENTS.....	4
ADCP	4
CALIBRATION.....	4
IMAGERY	4
DATA AND SCIENCE REPORT	4
XBT	4
XCTD.....	4
LOGSHEETS	4
SALTS.....	5
SCIENCE	5
MAPS.....	5
QC PLOTS	5
JGOFS DATA SET.....	6
TSG DATA FILES	6
RVDAS	8
<i>Meteorological and Light Data.....</i>	<i>8</i>
<i>Navigational Data</i>	<i>8</i>
<i>Geophysical Data</i>	<i>8</i>
<i>Oceanographic Data.....</i>	<i>8</i>
DATA FILE NAMES AND STRUCTURES.....	9
LKNU – KNUDSEN SONAR	9
LNDS – NET DEPTH SENSOR.....	10
LWN1 - WINCHES.....	10
LMWX - CAMPBELL METEOROLOGICAL DAS.....	10
UTSG – MICROTSG, THERMOSALINOGRAPH.....	11
LRTM – DIGITAL REMOTE TEMPERATURE	11
LDFLR – FLUOROMETER, WETLAB ECO	11
LOXY - OXYGEN	11
LPCO – PCO2 SYSTEM.....	12
LGUV – BIOSPHERICAL GUV	12
LSVP - SOUND VELOCITY PROBE IN ADCP TRANSDUCER WELL	13
LADC – ADCP SPEED LOG	13

LGYR - GYRO	14
LSEP – SEAPATH 330 GPS	14
LGAR - GARMIN GPS	16
TGPS – TRIMBLE CENTURION GPS	16
LMG SENSORS	19
SHIPBOARD SENSORS	19
CTD SENSORS	19
UNDERWAY CALIBRATION SHEETS	20
REMOTE TEMPERATURE (SBE-38)	20
THERMOSALINOGRAPH (SBE-45)	21
TRANSMISSOMETER (WETLABS C-STAR)	21
FLUOROMETER (WETLABS ECO-FL)	23
PSP (EPPLEY MODEL PSP)	24
PIR (EPPLEY MODEL PIR)	25
PAR (QSR-240P)	26
TEMPERATURE/RELATIVE HUMIDITY (RMYOUNG MODEL 41372LC)	27
BAROMETER (RMYOUNG MODEL 61201)	28
GUV (BIOSPHERICAL INSTRUMENTS 2511)	29
CTD FISH	30
PRIMARY TEMPERATURE	30
PRIMARY TEMPERATURE	31
SECONDARY TEMPERATURE	31
SECONDARY TEMPERATURE	32
PRIMARY CONDUCTIVITY	32
PRIMARY CONDUCTIVITY	33
SECONDARY CONDUCTIVITY	33
SECONDARY CONDUCTIVITY	34
PRIMARY DISSOLVED OXYGEN	34
PRIMARY DISSOLVED OXYGEN	35
.....	35
ACQUISITION AND PROCESSING INFORMATION	36
PROCESSING SPECIFICS	36
ERRORS AND EVENTS	36

Introduction

The LMG data acquisition systems continuously log data from a suite of instrumentation throughout the cruise. This document describes the format of that data and its location on the distribution CDs. It also contains important information that may affect how this data is processed such as instrument failures or other known problems with acquisition.

The data collected during this cruise is distributed on a CD-ROM written in ISO9660 level-1 format. This data format has very strict requirements on filenames and organization. However, it is readable by virtually every computing platform.

All of the data has been archived with the Unix “tar” command and/or compressed using Unix “gzip” compression. Tar files have a “.tar” extension and Gzipped files have a “.gz” extension. Tools are available on all platforms for uncompressing and de-archiving these formats. On Macintosh, Stuffit Expander with DropStuff will open a tar archive and uncompress gzipped and Unix compressed files. For Windows, WinZip, a shareware utility included on this CD (remember, it is shareware) will open these files.

In some cases to adhere to the ISO9660 format the .tar extension was removed. When we tarred the files then gzip the tar archive the name of the file became *File.tgz*. This name does not follow the 8.3 naming convention of the ISO9660 format. On Windows and Mac Platforms Winzip and Stuffit Expander handles this just fine. When they expand the *File.gz* the expanded file becomes *File.tar*, which both software packages can handle. On Unix platforms gunzip expands *File.tgz* but it does not append the .tar extension. So you may not recognize the file as a tar archive, but OS does recognize it as a tar archive. If you use the file command it will return saying it is a tar file. The below tar command will un-archive the file just fine.

IMPORTANT: Read the last section in this document, Acquisition Problems and Events, for important information that may affect the processing of this data.

Archive Data Extraction

It is often useful to know exactly how an archive was produced when expanding its contents. Tar files were created using the following commands:

```
tar cvf archive-file files-to-be-archived
```

To create a list of the files in the archive:

```
tar tvf archive-file > contents.list
```

To extract the files from the archive:

```
tar xvf archive-file file(s)-to-extract
```

G-zipped files will have a “.gz” extension on the filename. These files can be decompressed after de-archiving, using:

```
gunzip filename.gz
```

CD Directory Structure

<p>ADCP: ADCP.tar</p> <p>Cal: InstCoef.txt</p> <p>CTD: Ctd.tar</p> <p>Imagery: Imag.tar</p> <p>Maps:</p> <p>Process: JGOF.tar PCO2.tar PROC.tar QC.tar</p> <p>Report: Report.doc</p> <p>RVDAS: nav/ uw/</p> <p>Salinity: Logsheet.pdf Salts.xls</p> <p>Science: Utility: Acrobat Winzip</p> <p>XBT: Xbt.tar</p> <p>XCTD: Xctd.tar</p>		
---	--	--

Distribution Contents

ADCP

/Adcp/

This directory contains a tar file of gentoo's proc directory, which contains a database of the averaged ping data, Matlab m-files used in processing the data, and daily graphs of the currents. For more information contact Teri Chereskin at tchereskin@ucsd.edu.

Calibration

/Cal/

Refer to the InstCoef.txt file along with the specific instrument calibration sheets, both located in this directory, for information on how the RVDAS data was collected and processed.

Imagery

/Imagery/

This directory contains things such as ice imagery, isobar charts, sat imagery, wave and wind images, and weather reports.

Data and Science Report

/Report/

Copies of this report in MS Word, HTML, and text formats.

XBT

/XBT/

Contains a zip archive of XBT data generated for the Drake Transect by NOAA standard "AMVERSEAS" software. Non-Drake transect data may also be included, which will be a combination of binary and ascii files generated by standard Sippican MK-21 software.

XCTD

/XCTD/

Contains binary and ascii data files as generated by standard Sippican MK-21 software.

Logsheets

/logsheet/

This directory contains logsheets for XBT, XCTD, air samples and TCO2 data. For further information on this data (TCO2), contact Tim Newberger at tnewberg@ldeo.columbia.edu.

Salts

/Salinity/

This directory contains the log sheet for the Salt sample take during the Drake Transect Sampling, also a spreadsheet containing the Salt sample and TSG comparison. These samples were analyzed with the onboard AutoSal by the MST.

Science

/Science/

This directory, if populated, contains data specified by the on-board science party.

Maps

/Maps/

This directory maps and mapping data generated by the ship's MCIS and automated scripts, usually in JPEG or PostScript format, plus any maps provided for this purpose by the on-board science party.

QC Plots

/QC_PLOTS/

Postscript files of data stored each day on RVDAS for quality control analysis during the cruise. There are 3 types of files, named metXXX.ps, navXXX.ps, and oceanXXX.ps, where XXX represents the Julian day. Met files are a summary of the data from the meteorological instruments, Nav files are a summary of navigational data, and Ocean files are a summary of the underway seawater and bathymetry data.

JGOFS Data Set

/JGOF/

The JGOFS data set consists of a single file produced each day named jg<julian_day>.dat.gz where <julian_day> is the day the data was acquired. The “.gz” extension indicates that the individual files are compressed before archiving. The daily file consists of 22 separate columnar fields in text format, which are described below. The JGOFS data set is obtained primarily by applying calibrations to raw data and decimating to whole minute intervals. However, several fields are derived measurements from more than a single raw input. *Note: Null, unused, or unknown fields are filled with 9's in the JGOFS data.*

Additionally, 3 separate QC plots are generated daily by the ET using the JGOFS data set. These plots include TSG and Bathymetry data, meteorological data, and navigation data. The files are called ocean<julian_day>.ps, met<julian_day>.ps, and nav<julian_day>.ps respectively.

Field	Data	Units
01	GMT date	dd/mm/yy
02	GMT time	hh:mm:ss
03	PCOD latitude (negative is South)	Ddd.dddd
04	PCOD longitude (negative is West)	Ddd.dddd
05	Ships speed	Knots
06	GPS HDOP	-
07	Gyro Heading	Degrees (azimuth)
08	Course over ground	Degrees (azimuth)
09	Mast PAR	$\mu\text{Einsteins/meters}^2 \text{ sec}$
10	Sea surface temperature	°C
11	Not used	-
12	Sea surface salinity	PSU
13	Sea depth (uncorrected, calc. sw sound vel. 1500 m/s)	meters
14	True wind speed (port windbird)	meters/sec
15	True wind direction (port windbird)	degrees (azimuth)
16	Ambient air temperature	°C
17	Relative humidity	%
18	Barometric pressure	mBars
19	Sea surface fluorometry	volts (0-5 FSO)
20	Not used	-
21	PSP	W/m^2
22	PIR	W/m^2

TSG Data files

/TSG/tsgfl

RVDAS processes the ltsg.d### file, using the Seabird calibration. It produces a daily tsgfl.d### file with the below fields.

04+321:00:01:23.978 -00.070 -00.089 02.8042 33.75690 0.471306 4.341880

Field	Data	Units
1	RVDAS Time Tag	
2	Internal water temperature	°C
3	External water temperature	°C
4	Conductivity	S/cm
5	Salinity	PSU

Field	Data	Units
6	Transmissometer signal	Volts

RVDAS

/RVDAS/

RVDAS (Research Vessel Data Acquisition System) was developed at Lamont-Doherty Earth Observatory of Columbia University and has been used on the R/V Maurice Ewing for several years. It was adapted for use on the Nathaniel B. Palmer and her sister ship, the R/V Laurence M. Gould.

Below you will find detailed information on the data included. Be sure to read the “Significant Acquisition Events” section below for important information about data acquisition during this cruise.

Meteorological and Light Data

Measurement	File ID	Collect. Status	Rate	Instrument
Air Temperature	lmwx	continuous	1 sec	R. M. young 41372VC
Relative Humidity	lmwx	continuous	1 sec	R. M. young 41372VC
Wind Speed/Direction	lmwx	continuous	1 sec	Gill Ultrasonic
PAR, (Photosynthetically-Available Radiation)	lmwx	continuous	1 sec	BSI QSR-240
Barometer	lmwx	continuous	1 sec	R. M. young 61201
GUV	lguv	continuous	1 sec	GUV2511
PIR (LW radiation)	lmwx	continuous	1 sec	Eppler PIR
PSP (SW radiation)	lmwx	continuous	1 sec	Eppler PSP
Port Ultrasonic Wind Speed/Direction	lmwx	continuous	1 sec	Gill Wind Observer II

Navigational Data

Measurement	File ID	Collect. Status	Rate	Instrument
Trimble GPS	tgps	civilian mode	1 sec	Trimble 20636-00SM
Gyro	lgyr	continuous	0.2 sec	Anschutz Gyro
Garmin GPS	lgar	continuous	1 sec	Garmin 17
Seapath GPS	Lsep	Continuous	1 sec	Seapath 330

Geophysical Data

Measurement	File ID	Collect. Status	Rate	Instrument
Bathymetry	lknu	variable	Varies	Knudsen 320B/R
DUSH 11 Winch	lwn1	variable	varies	Markey DUSH 11
DUSH 5 Winch	lwn1	variable	varies	Markey DUSH 5
DUSH 4 Winch	lwn1	variable	varies	Markey DUSH 4

Oceanographic Data

Measurement	File ID	Collect. Status	Rate	Instrument
Salinity	utsg	continuous	1 sec	SeaBird 45
Sea Surface Temperature	Lrtm	continuous	1 sec	SeaBird 38
Fluorometry (digital)	lfir	continuous	1 sec	Turner 10-AU-005
Fluorometry (digital)	ldflr	continuous	1 sec	Wetlab ECO
ADCP, Speed Log	ladc	continuous	1 sec	RD Instruments
Oxygen	loxy	continuous	1 sec	
PCO2	lpco2	continuous	2.5 min	

Data File Names and Structures

RVDAS data is divided into two broad categories, **Underway** and **Navigation**. The groups are abbreviated “uw” and “nav”. Thus, these two tar files, lmguw.tar and lmgnav.tar exist under the top-level rvdas directory. The instruments are broken down as shown. Each data file is g-zipped to save space on the distribution. Not all data types are collected everyday or on every cruise.

RVDAS data files are named following the convention: LMG[FileID].dDDD.

- The FileID is a 4-character code representing the system being logged, for example: lmet (for meteorology)
- DDD is the Julian day of the data collection

Underway Data	File ID	Navigation Data	File ID
Meteorological - Cambell	lmwx	Gyro Compass	lgyr
Knudsen	lknu	P-CODE GPS	tgps
microTSG	utsg	Garmin 17 GPS	lgar
Digital Remote Temperature	lrtm	Seapath 330 GPS	lsep
Fluorometer – Wetlab ECO	ldfl		
ADCP	ladc		
Sound Velocity Probe	lsvp		
GUV & PUV	lguv		
PCO2 System	lpco		
Oxygen	loxy		
Wet Wall Flows	lsea		
Winches: Dush4,5,&11	lwn1		
Net Depth Sensor	lnds		

Data is received by the RVDAS system via RS-232 serial connections. The data files that comprise the rvdas data set are described below. A time tag is added to each line of data received and the data is written to disk.

YY+DDD:HH:MM:SS.SSS [data stream from instrument]

Where, YY: two-digit year, DDD: Julian Day, HH: 2 digit hours, MM: 2 digit minutes SS.SSS: seconds. All times are UTC.

The delimiters used to separate fields in the raw data files are usually spaces and commas, but other delimiters are used (:, =, @) and occasionally there is no delimiter. Care should be taken when reprocessing the data that the fields separations are clearly understood. An example data

lknu – Knudsen Sonar

08+024:07:36:36.245 HF,00.00, 000,0,LF,448.9,-026,1

Field	Data	Units
1	RVDAS Time Tag	
2	HF – high frequency header (12 kHz)	
3	HF - depth to surface	meters
4	HF - Echo Strength	
5	HF – Depth Valid Flag	
6	LF – low frequency header (3.5 kHz)	
7	LF - depth to surface	meters
8	LF - Echo Strength	
9	LF – Depth Valid Flag	

Inds – Net Depth Sensor

99+099:00:18:19.775 V01 00199.8

Field	Data	Units
1	RVDAS Time Tag	
2	V01 – Sensor 1	label
3	Depth	meters

lwn1 - Winches

08+033:11:27:50.673 RD,DUSH-5,00111.63,00000000,-0000012,1938

Field	Data	Units
1	RVDAS Time Tag	
2	Record Identifier, RD=Remote Data	
3	Winch Identifier, DUSH-X where X is 4, 5, or 11	
4	Tension	lbs
5	Speed	Meters/minute
6	Payout	meters
7	Checksum	

Imwx - Campbell Meteorological DAS

08+034:13:52:14.216 PUS,A,356,002.15,M,+332.28,+000.97,60,08

Field	Data	Units
1	RVDAS Time Tag	
2	PUS tag – Port UltraSonic Anemometer	
3	Unit Identification, A-Z	
4	Port Wind Direction, degrees relative to Bow	deg
5	Port Wind Speed	m/s
6	Units, M=meters per second	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Status, 0=ok, 60=Heating Enabled & ok, Other value mean a fault	
10	Check Sum	

08+034:13:52:14.216 SUS,A,356,002.15,M,+332.28,+000.97,60,08

Field	Data	Units
1	RVDAS Time Tag	
2	SUS tag – Starboard UltraSonic Anemometer	
3	Unit Identification, A-Z	
4	Port Wind Direction, degrees relative to Bow	deg
5	Port Wind Speed	m/s
6	Units, M=meters per second	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Status, 0=ok, 60=Heating Enabled & ok, Other value mean a fault	
10	Check Sum	

08+034:13:52:14.454 MET,12.22322,44.25706,-75,-25,-363.6365,2.332982,-0.08215196,278.6845,279.2192,854.6198

Field	Data	Units
1	RVDAS Time Tag	
2	MET tag	
3	Power Supply Voltage	Volts
4	Enclosure Relative Humidity	%
5	Air Temp	°C
6	Air Relative Humidity	%
7	PAR	mVolts
8	PSP Thermopile	mVolts
9	PIR Thermopile	mVolts
10	PIR Case Temperature	°K
11	PIR Dome Temperature	°K
12	Barometer	mBars

utsg – microTSG, Thermosalinograph

For further information on this data, check www.seabird.com for SBE 45 MicroTSG Thermosalinograph

08+037:13:45:57.596 2.6470, 3.03853, 33.8129, 1459.351

Field	Data	Units
1	RVDAS Time Tag	
2	Internal water temperature	°C
3	Conductivity	S/m
4	Salinity	psu
5	Sound Velocity	m/s

Irtm – digital Remote Temperature

For further information on this data, check on www.seabird.com on SBE38 Digital Thermometer

08+037:13:47:17.841 2.2527

Field	Data	Units
1	RVDAS Time Tag	
2	External water temperature	°C

Idflr – Fluorometer, Wetlab ECO

08+037:13:55:08.434 99/99/99 99:99:99 0.00 2585 73 543

Field	Data	Units
1	RVDAS Time Tag	
2	Fluorometer Date	mm/dd/yy
3	Fluorometer Time	hh:mm:ss
4	Chlorophyll Signal	µg/l
5	Reference	λq
6	Counts – Chlorophyll Signal	Count
7	Thermistor	

loxy - Oxygen

For further information on this data, contact Tim Newberger at tnewberg@ldeo.columbia.edu

04+117:23:57:23.504 MEASUREMENT 3830 380 Oxygen: 309.95 Saturation:
83.48 Temperature: -1.35 DPhase: 33.41 BPhase: 32.22

0.00 RPhase: 0.00 Bamp: 262.09 BPot: 163.00 Ramp:
RawTem.: 694.92

Field	Data	Units
1	RVDAS Time Tag	
2-4	Measurement ID, Model Number, Serial Number	alphanumeric
5	Oxygen heading	text
6	Oxygen Reading	Raw numeric
7	Saturation heading	text
8	Saturation Reading	Raw numeric
9	Temperature heading	text
10	Water Temperature	°C
11	Dphase heading	text
12	Dphase	Raw numeric
13	Bphase heading	text
14	BPhase	Raw numeric
15	Rphase heading	text
16	Rphase	Raw numeric
17	Bamp heading	text
18	Bamp	Raw numeric
19	Bpot heading	text
20	Bpot	Raw numeric
21	Ramp heading	text
22	Ramp	Raw numeric
23	RawTem heading	text
24	RawTemp	Raw numeric

IpcO – PCO₂ system

For further information on this data, contact Tim Newberger at tnewberg@ldeo.columbia.edu

02+319:23:59:13.748 2002319.99851 7154.27 26.49 1033.6 325.79 6.74 329.3
53.76 0 Equil

Field	Data	Units
1	RVDAS Time Tag	
2	Julian date file string	Julian
3	IR voltage reading	mV
4	Cell temperature	°C
5	Barometer	millibars
6	VCO ₂	mL
7	Equilibrator temperature	°C
8	PCO ₂	millibars
9	Gas flow	mL/min
10	Solenoid position ID	number
11	Valve Position ID	number
12	Measured gas	name

LgUV – Biospherical GUV

08+037:14:17:59.211 020608 141758 -.000099 1.307E0 7.24E0 1.316E1 2.609E1 3.285E1 3.505E1 8.075E-
2 38.993 17.985

GUV only

Field	Data	Units
1	RVDAS Time Tag	
2	GUV Computer Date	mmddyy

GUV and PUV

Field	Data	Units
1	RVDAS Time Tag	
2	GUV Computer Date	mmddyy

3	GUV Computer Time	hhmmss
4	Ed0Gnd - GUV	Volts
5	Ed0305 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
6	Ed0313 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
7	Ed0320 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
8	Ed0340 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
9	Ed0380 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
10	Ed0395 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
11	Ed0PAR - GUV	$\mu\text{E}/\text{cm}^2\text{sec}$
12	Ed0Temp - GUV	$^{\circ}\text{C}$
13	Ed0VIn	Volts

3	GUV Computer Time	hhmmss
4	EdZGnd -PUV	Volts
5	EdZ305 -PUV	$\mu\text{W}/\text{cm}^2\text{nm}$
6	EdZ313 -PUV	$\mu\text{W}/\text{cm}^2\text{nm}$
7	EdZ320 -PUV	$\mu\text{W}/\text{cm}^2\text{nm}$
8	EdZ395 -PUV	$\mu\text{W}/\text{cm}^2\text{nm}$
9	EdZ340 -PUV	$\mu\text{W}/\text{cm}^2\text{nm}$
10	EdZPAR -PUV	$\mu\text{E}/\text{cm}^2\text{sec}$
11	LuZChl -PUV	$\mu\text{E}/\text{srm}^2\text{sec}$
12	EdZ380 -PUV	$\mu\text{W}/\text{cm}^2\text{nm}$
13	WTemp -PUV	$^{\circ}\text{C}$
14	Depth -PUV	m
15	EdZTemp -PUV	$^{\circ}\text{C}$
16	LuZTemp -PUV	$^{\circ}\text{C}$
17	Tilt -PUV	Degrees
18	Roll -PUV	Degrees
19	Ed0Gnd - GUV	Volts
20	Ed0305 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
21	Ed0313 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
22	Ed0320 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
23	Ed0340 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
24	Ed0380 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
25	Ed0395 - GUV	$\mu\text{W}/\text{cm}^2\text{nm}$
26	Ed0PAR - GUV	$\mu\text{E}/\text{cm}^2\text{sec}$
27	Ed0Temp - GUV	$^{\circ}\text{C}$
28	Ed0VIn	Volts

Isvp - Sound Velocity Probe in ADCP Transducer Well

00+348:01:59:52.128 1539.40

Field	Data	Units
1	RVDAS Time Tag	
2	Sound velocity	m/s

ladc – ADCP Speed Log

00+019:23:59:59.099 \$PUHAW,UVH,-1.48,-0.51,250.6

Field	Data	Units
1	RVDAS Time Tag	
2	\$PUHAW	
3	UVH (E-W, N-S, Heading)	
4	Ship Speed relative to reference layer ¹ velocity ² , East vector	knots
5	Ship Speed relative to reference layer ¹ velocity ² , North vector	knots
6	Ship heading	degrees

¹The reference layer is an average velocity measured in a number of depth “bins”. On the LMG, the bins are eight meters deep and bins 3-10 define the reference layer. Hence, the reference layer is the water column from 16-80 meters beneath the ship.

²The speed output is water velocity relative to the ship’s hull and is therefore opposite of the actual movement of the ship. For example, if the ship’s heading is due north, the North/South reference layer velocity is likely to be negative (southerly).

Igyr - Gyro

02+315:23:59:58.194 \$PASVW,00.1,A*1D

02+315:23:59:58.414 \$IIVHW,287.7,T,,M,,N,,K*71

02+315:23:59:58.616 \$HEHDT,287.7,T*25

02+315:23:59:58.821 \$HEROT,001.6,A*2C

02+315:23:59:58.984 \$HCHDT,,T*07

HDT: True Heading

01+083:00:00:02.893 \$HEHDT,246.3,T*2C

Field	Data	Units
1	RVDAS Time Tag \$HEHDT	
2	Heading XXXXX = ddd.d	degrees
3	T flag for true heading, checksum	

ROT: Rate of Turn

01+083:00:00:03.093 \$HEROT,-006.3,A*03

Field	Data	Units
1	RVDAS Time Tag \$HEROT	
2	Rate of turn	degrees/min
3	Status: A = data valid, checksum	

Isep – Seapath 330 GPS**INZDA: Time and Date Data**

10+351:23:59:58.142 \$INZDA,235958.08,17,12,2010,,*78

Field	Data	Units
1	RVDAS Time Tag	
2	\$INZDA Tag	
3	time	hhmmss.ss
3	day	dd
4	month	mm
5	year	yyyy
6	(blank)	
7	Check sum	hexadecimal

INGGA: Global Positioning Fix Data

10+351:23:59:58.142 \$INGGA,235958.07,6118.168460,S,06008.089527,W,1,12,0.7,22.57,M,17.79,M,,*46

Field	Data	Units
1	RVDAS Time Tag	
2	\$INGGA Tag	
3	Time (UTC)	hhmmss.ss
3	Latitude in degrees with decimal minutes	ddmm.mmm
4	{NIS} (latitude is north or south)	
5	Longitude in degrees with decimal minutes	ddmm.mmm
6	{EIW} (longitude is east of west)	
7	GPS quality indicator: 0 = invalid position, 1 = GPS SPS used, 2 = DGPS used, 3 = GPS PPS used, 4 = GPS RTK used, 5 = GPS float RTK used, 6 = dead reckoning	
8	Number of Satellites in use (00-99)	
9	HDOP	

Field	Data	Units
10	Height above ellipsoid in meters	m.mm
11	M	
12	Age of DGPS corrections in seconds	ss.ss
13	M	
14	(blank)	
15	*Check sum	hexadecimal

INRMC: Recommended Minimum Specific GNSS Data

10+351:23:59:58.200 \$INRMC,235958.07,A,6118.168460,S,06008.089527,W,12.8,331.22,171210,11.3,E,A*1C

Field	Data	Units
1	RVDAS Time Tag	
2	\$INRMC Tag	
3	UTC of position	hhmmss.ss
4	Status A = Data Valid, V = Navigation Receiver Warning	
5	Latitude in degrees with decimal minutes	ddmm.mmm
6	North (N) or South (S)	
7	Longitude in degrees with decimal minutes	ddmm.mmm
8	East (E) or West (W)	
9	Speed Over Ground, knots	knots
10	Course Over Ground, degrees True	degrees
11	Date	ddmmyy
12	Magnetic Variation, degrees E/W	degrees
13	Mode Indicator E= Estimated Mode	
14	*Check sum	

PSXN,20: Data Quality

10+351:23:59:58.200 \$PSXN,20,1,2,0,0*38

Field	Data	Units
1	RVDAS Time Tag	
2	\$PSXN Tag	
3	20 (PSXN identifier)	
3	Horizontal position and velocity quality: 0 = normal, 1 = reduced performance, 2 = invalid data	
4	Height and vertical velocity quality: 0 = normal, 1 = reduced performance, 2 = invalid data	
5	Heading quality: 0 = normal, 1 = reduced performance, 2 = invalid data	
6	Roll and pitch quality: 0 = normal, 1 = reduced performance, 2 = invalid data	
7	*Check sum	hexadecimal

PSXN,23: Roll, Pitch, Heading and Heave

10+351:23:59:58.213 \$PSXN,23,0.02,-0.76,330.56,*0B

Field	Data	Units
1	RVDAS Time Tag	
2	\$PSXN Tag	
3	23 (PSXN identifier)	
3	Roll in degrees. Positive with port side up.	d.dd
4	Pitch in degrees. Positive with bow up.	d.dd
5	Heading in degrees true	d.dd
6	Heave in meters. Positive is down	m.mm
7	*Check sum	hexadecimal

Igar - Garmin GPS

GGA: Global Positioning Fix Data

08+034:12:26:06.131 \$GPGGA,122607,6446.4733,S,06403.4455,W,1,11,0.9,-193.4,M,9.7,M,,*5A

Field	Data	Units
1	RVDAS Time Tag	
2	\$GPGGA Tag	
3	UTC of position	hhmmss.ss
4	Latitude in degrees with decimal minutes	ddmm.mmm
5	North (N) or South (S)	
6	Longitude in degrees with decimal minutes	ddmm.mmm
7	East (E) or West (W)	
8	GPS quality (1=GPS 2=DGPS)	
9	Number of GPS satellites used	
10	Horizontal dilution of precision (HDOP)	
11	Antenna height above/below mean-sea-level (geoid)	meters
12	Units for antenna height (M = Meters)	
13	Geoidal Separation ¹	
14	Units for Geoidal Separation (M = Meters)	meters
15	Age of differential GPS data, number of seconds since last SC104 Type 1 or 9	
16	Differential reference station ID	

¹Geoidal Separation: the difference between the WGS-84 earth ellipsoid and mean-sea-level (geoid). A negative value represents mean-sea-level below ellipsoid.

GLL: Geographic Position – Latitude/Logitude

08+034:12:26:06.211 \$GPGLL,6446.4733,S,06403.4455,W,122607,A

Field	Data	Units
1	RVDAS Time Tag	
2	\$GPGLL Tag	
3	Latitude	ddmm.mmm
4	North (N) or South (S)	
5	Logitude	ddmm.mmm
6	East (E) or West (W)	
7	UTC of position	hhmmss.ss
8	Status: A = Data Valid, V =Data Not Valid	

VTG: Track Made Good and Speed over Ground

08+034:12:26:06.211 \$GPVTG,167,T,151,M,000.0,N,0000.0,K

Field	Data	Units
1	RVDAS Time Tag	
2	\$GPVTG Tag	
3	Track, degrees true	degrees
3	T flag for True	
4	Track, degrees magnetic	degrees
5	M flag for Magnetic	
6	Speed over Ground	knots
7	N flag for Knots	
8	Speed over Ground	kmhr
9	K flag for km/hr	

tgps – Trimble Centurion GPS

GGA: Global Positioning Fix Data

08+034:12:26:06.131 \$GPGGA,122607,6446.4733,S,06403.4455,W,1,11,0.9,-193.4,M,9.7,M,,*5A

Field	Data	Units
1	RVDAS Time Tag	
2	\$GPGGA Tag	
3	UTC of position	hhmmss.ss
4	Latitude in degrees with decimal minutes	ddmm.mmm
5	North (N) or South (S)	
6	Longitude in degrees with decimal minutes	ddmm.mmm
7	East (E) or West (W)	
8	GPS quality (1=GPS 2=DGPS)	
9	Number of GPS satellites used	
10	Horizontal dilution of precision (HDOP)	
11	Antenna height above/below mean-sea-level (geoid)	meters
12	Units for antenna height (M = Meters)	
13	Geoidal Separation ¹	
14	Units for Geoidal Separation (M = Meters)	meters
15	Age of differential GPS data, number of seconds since last SC104 Type 1 or 9	
16	Differential reference station ID	

¹Geoidal Separation: the difference between the WGS-84 earth ellipsoid and mean-sea-level (geoid). A negative value represents mean-sea-level below ellipsoid.

GLL: Geographic Position – Latitude/Longitude

08+034:12:26:06.211 \$GPGLL,6446.4733,S,06403.4455,W,122607,A

Field	Data	Units
1	RVDAS Time Tag	
2	\$GPGLL Tag	
3	Latitude	ddmm.mmm
4	North (N) or South (S)	
5	Longitude	ddmm.mmm
6	East (E) or West (W)	
7	UTC of position	hhmmss.ss
8	Status: A = Data Valid, V = Data Not Valid	

VTG: Track Made Good and Speed over Ground

08+034:12:26:06.211 \$GPVTG,167,T,151,M,000.0,N,0000.0,K

Field	Data	Units
1	RVDAS Time Tag	
2	\$GPVTG Tag	
3	Track, degrees true	degrees
3	T flag for True	
4	Track, degrees magnetic	degrees
5	M flag for Magnetic	
6	Speed over Ground	knots
7	N flag for Knots	
8	Speed over Ground	kmhr
9	K flag for km/hr	

RMC: Recommended Minimum Specific GNSS Data

08+034:13:17:26.627 \$GPRMC,131726.605,A,6446.4820,S,06403.3075,W,000.0,094.4,030208,16.3,E

Field	Data	Units
1	RVDAS Time Tag	
2	\$GPRMC Tag	

Field	Data	Units
3	UTC of position	hhmmss.ss
4	Status A = Data Valid, V = Navigation Receiver Warning	
5	Latitude in degrees with decimal minutes	ddmm.mmm
6	North (N) or South (S)	
7	Longitude in degrees with decimal minutes	ddmm.mmm
8	East (E) or West (W)	
9	Speed Over Ground, knots	knots
10	Course Over Ground, degrees True	degrees
11	Date	ddmmyy
12	Magnetic Variation, degrees E/W	degrees
13	Mode Indicator E= Estimated Mode	

LMG Sensors

Shipboard Sensors

Sensor	Description	Serial #	Cal. Date	Status
Port Anemometer	Gill Ultrasonic Wind Observer II	840019	N/A	Collected
Starboard Anemometer	Gill Ultrasonic Wind Observer II	71738	N/A	Collected
Barometer	R.M. Young 61201	BP0873	28-Jul-2010	Collected
Humidity/Wet Temp	RM Young 41372LC	06720	11-Feb-2010	Collected
PAR for Mast	Biosph. Inst. QSR-240P	6393	31-Aug-2010	Collected
PIR	Eppley PIR	32031F3	15-Jan-2010	Collected
PSP	Eppley PSP	31701F3	15-Jan-2010	Collected
GUV (Mast)	Biosph. Inst. GUV-2511	25110805127	17-May-2010	Collected
Transmissometer	WET Labs C-Star 25 cm deep	CST-553DR	26-Aug-2010	Collected
MicroTSG	Sea-Bird 45	243	21-Jul-2010	Collected
Digital Remote Temp	Sea-Bird 38	324	26-Jan-2010	Collected
Fluorometer	WET Labs ECO-FL	FLRTD-399	15-Sep-09	Collected

CTD Sensors

Sensor	Description	Serial #	Cal. Date	Status
CTD Fish	Seabird SBE9Plus	0377	22-Feb-11	Collected
Primary Temperature	Seabird SBE3	1080	11-Aug-09	Collected
Secondary Temperature	Seabird SBE3	1085	11-Aug-09	Collected
Primary Conductivity	Seabird SBE4	3648	11-Sep-09	Collected
Secondary Conductivity	Seabird SBE4	0763	11-Aug-09	Collected
Primary Dissolved Oxygen	Seabird SBE43	430196	29-Sept-10	Collected

Underway Calibration Sheets

Note: Embedded pdf files can be opened with a right-click, Acrobat Document Object, Open

Remote Temperature (SBE-38)

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0324
CALIBRATION DATE: 26-Jan-10

SBE 38 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE


ITS-90 COEFFICIENTS

a0 = -2.126219e-005
a1 = 2.756564e-004
a2 = -2.490794e-006
a3 = 1.524467e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.50010	895419.1	-1.50002	0.00008
0.99990	798448.4	0.99984	-0.00006
4.50000	682058.1	4.49983	-0.00017
7.99990	584568.4	7.99997	0.00007
11.49990	502633.3	11.50003	0.00013
15.00000	433539.8	15.00002	0.00002
18.50000	375083.6	18.50001	0.00001
22.00000	325471.2	21.99994	-0.00006
25.50000	283232.8	25.49995	-0.00005
30.00000	237850.6	29.99998	-0.00002
32.50000	216277.2	32.50005	0.00005

Temperature ITS-90 = $1/(a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]) - 273.15$ (°C)

Residual = instrument temperature - bath temperature



Thermosalinograph (SBE-45)

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0243
CALIBRATION DATE: 21-Jul-10SBE 45 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.008835e+000
h = 1.573683e-001
i = -3.239483e-004
j = 5.119501e-005CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 1.1173e-006


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2535.87	0.00000	0.00000
1.0000	34.9328	2.98497	5043.04	2.98496	-0.00001
4.4999	34.9132	3.29295	5233.27	3.29297	0.00001
15.0000	34.8700	4.27750	5798.83	4.27750	-0.00000
18.5000	34.8603	4.62357	5984.66	4.62357	-0.00000
23.9999	34.8493	5.18298	6273.11	5.18299	0.00001
29.0000	34.8428	5.70617	6531.01	5.70615	-0.00002
32.5000	34.8381	6.07936	6708.74	6.07937	0.00001

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure(decibars); δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity



Transmissometer (Wetlabs C-Star)

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration

Date **August 26, 2010**

S/N#

CST-553DR

Pathlength **25 cm**

	Analog meter
V_d	0.058 V
V_{air}	4.782 V
V_{ref}	4.687 V

Temperature of calibration water **24.9 °C**

Ambient temperature during calibration **23.6 °C**

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x , in meters): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x * \ln(Tr)$

V_d Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref} .

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.

Revision L

6/9/09

Fluorometer (Wetlabs ECO-FL)

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5850
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 9/15/2009

S/N: FLRTD-399

Chlorophyll concentration expressed in µg/l can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

	Analog Range 1	Analog Range 2	Analog Range 4 (default)	Digital
Dark Counts	0.097	0.054	0.033 V	68 counts
Scale Factor (SF)	6	12	24 µg/l/V	0.0074 µg/l/count
Maximum Output	4.96	4.96	4.96 V	16330 counts
Resolution	0.7	0.7	0.7 mV	1.0 counts

Ambient temperature during characterization

22.3 °C

Analog Range: 1 (most sensitive, 0–4,000 counts), 2 (midrange, 0–8,000 counts), 4 (entire range, 0–16,000 counts).

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $SF = x / (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations *in-situ* is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

FLRTD-399_workbookj1.xls

Revision J

3/17/08

PSP (Eppley model PSP)**THE EPPLEY LABORATORY, INC.**

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840
Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

**STANDARDIZATION OF
EPPLEY PRECISION SPECTRAL PYRANOMETER
Model PSP**

Serial Number: 31701F3

Resistance: 674 Ω at 23°C

Temperature Compensation Range: -20° to + 40°C

This radiometer has been compared with Standard Precision Spectral Pyranometer, Serial Number 21231F3 in Eppley's Integrating Hemisphere under radiation intensities of approximately 700 watts meter⁻² (roughly one half a solar constant).

As a result of a series of comparisons, it has been found to have a sensitivity of:

$$8.36 \times 10^{-6} \text{ volts/watts meter}^{-2}$$

The calculation of this constant is based on the fact that the relationship between radiation intensity and emf is rectilinear to intensities of 1400 watts meter⁻². This radiometer is linear to within $\pm 0.5\%$ up to this intensity.

The calibration of this instrument is traceable to standard self-calibrating cavity pyrheliometers in terms of the Systems Internationale des Unites (SI units), which participated in the Tenth International Pyrheliometric Comparisons (IPC X) at Davos, Switzerland in September-October 2005.

Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy. Unless otherwise stated in the remarks section below or on the Sales Order, the results are "AS FOUND / AS LEFT".

Useful conversion facts: 1 cal cm⁻² min⁻¹ = 697.3 watts meter⁻²
1 BTU/ft²-hr⁻¹ = 3.153 watts meter⁻²

Shipped to: Raytheon Technical Services (NSF) Date of Test: January 15, 2010
Port Hueneme, CA

S.O. Number: 62361
Date: January 20, 2010

In Charge of Test: *Dale L. Bortz*
Reviewed by: *Thomas D. Smith*

Remarks:

PIR (Eppley model PIR)**THE EPPLEY LABORATORY, INC.**

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840
Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

**STANDARDIZATION OF
EPPLEY PRECISION INFRARED RADIOMETER
Model PIR**

Serial Number: 32031F3

Resistance: 709 Ω at 23°C

Temperature Compensation Range: -20° to + 40°C

This pyrgeometer has been compared against Eppley's Blackbody Calibration System under radiation intensities of approximately 200 watts meter⁻² and an average ambient temperature of 23°C as measured by Standard Omega Temperature Probe, RTD#1.

As a result of a series of comparisons, it has been found to have a sensitivity of:

$$3.86 \times 10^{-6} \text{ volts/watts meter}^{-2}$$

The calculation of this constant is based on the fact that the relationship between radiation intensity and cmf is rectilinear to intensities of 700 watts meter⁻². This radiometer is linear to within $\pm 1.0\%$ up to this intensity.

The calibration of this instrument is traceable to the International Practical Temperature Scale (IPTS) through a precision low-temperature blackbody.

Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy. Unless otherwise stated in the remarks section below or on the Sales Order, the results are "AS FOUND / AS LEFT".

Shipped to: Raytheon Technical Services (NSF) Date of Test: January 15, 2010
Port Hueneme, CA

S.O. Number: 62360
Date: January 20, 2010

In Charge of Test: *Colin L. Smith*

Reviewed by: *Thomas D. Kish*

Remarks:

PAR (QSR-240P)**Biospherical Instruments Inc.****CALIBRATION CERTIFICATE**

Calibration Date: 8/31/2010
 Model Number: QSR240
 Serial Number: 6393
 Operator: TPC
 Standard Lamp: GS-1024(B/25/08)
 Probe Excitation Voltage Range: 6 to 10 VDC(+)

Output Polarity: Positive

Probe Conditions at Calibration (in air):

Calibration Voltage: 6 VDC(+)

Probe Current: 1.3 mA

Probe Output Voltage:

Probe Illuminated: 95.9 mV
 Probe Dark: 0.3 mV
 Probe Net Response: 95.6 mV
 RG780: 0.4 mV

Corrected Lamp Output:

Output in Air (same condition as calibration):

9.271E+15 quanta/cm²sec
0.01540 μE/cm²sec

Calibration Scale Factor:

(To calculate irradiance, divide the net voltage reading in Volts by this value.)

Dry: 1.0310E-17 V/(quanta/cm²sec)
6.2087E+00 V/(μE/cm²sec)

Notes:

1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).
3. The collector should be cleaned frequently with alcohol.
4. Calibration was performed with customer cable, when available.

QSR240R 03/24/95

Temperature/Relative Humidity (RMYoung model 41372LC)

Meteorological Instruments

COPY

Temperature Sensor Calibration ReportCustomer: *Raytheon Technical Services Co*

Test Number: 44159

Customer PO: RM10889-50

Test Date: 15 April 2004

Sales Order: 7108

Test Sensor:

Model: 41372LC

Serial Number: 6720

Description: Temperature/Relative Humidity Sensor

Report of calibration comparison of test temperature sensor with National Institute of Standards and Technology traceable standard thermometers at three temperatures in the R.M. Young Company controlled temperature calibration bath facilities. Calibration accuracy $\pm 0.1^\circ$ Celsius.

Bath Temperature (degrees C)	Current Output (milliamps)	Indicated (1) Temperature (degrees C)
-49.98	4.011	-49.93
0.03	12.006	0.04
49.97	19.994	49.96

(1) Calculated from current output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference InstrumentSerial # NIST Test Reference

Brooklyn Thermometer Model 43-FC

8006-118

204365

Brooklyn Thermometer Model 22332-D5-FC

25071

249763

Brooklyn Thermometer Model 2X400-D7-FC

77532

228060

Keithley Multimeter Model 191

15232

234027

Tested By: 

R.M. YOUNG COMPANY 2801 Aero Park Drive, Traverse City, Michigan 49686 USA

Tel: 231-946-3980 Fax: 231-946-4772 Email: met.sales@youngusa.com

Barometer (RMYoung model 61201)

R.M. Young Company
2901 Aero Park Drive
Traverse City, Michigan 49686 USA



CALIBRATION REPORT
Barometric Pressure Sensor

Customer: *Raytheon Technical Services Company LLC*

Test Number: 07261

Customer ID: RR52837-01

Test Date: 28 July 2010

Sales Order: 1325

Test Sensor:

Model: 61201

Serial Number: BP00873

Description: Barometric Pressure Sensor

Report of calibration comparison of test barometric pressure sensor with National Institute of Standards and Technology traceable standard pressure calibrator at five pressures in the R.M. Young Company controlled pressure facility. Calibration accuracy ± 1.0 hPa.

Reference Pressure (hPa)	Voltage Output (millivolts)	Indicated (1) Pressure (hPa)
800.0	0	800.0
875.0	1251	875.1
950.0	2501	950.1
1025.0	3750	1025.0
1100.0	4997	1099.8

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference Instrument

Druck Pressure Controller Model DPI515
Fluke Multimeter Model 8080A

Serial NIST Test Reference

5150041 UKAS Lab 0221
486546 234027

Tested By:

E. Channing

METEOROLOGICAL INSTRUMENTS

Tel: 231-646-3990 Fax: 231-646-4772 Email: met.sales@youngusa.com Web: youngusa.com

GUV (Biospherical Instruments 2511)



Biospherical Instruments Inc.

GUV-2511 Calibration Certificate

System Serial Number		25110005127		Date of Calibration		5/17/2010										
Calibration database		25110005127.v4.mdb		Date of Certificate		5/17/2010										
DASSN		0111		Standard of Spectral Irradiance		G81018(A28108)										
Microprocessor Tag Number		2		Operator		TC										
Monochromator Channels	Address	Wavelength [nm]	Responsivity			Offset Small [pA]	Offset Medium [pA]	Offset Large [pA]	Measurement Units							
			Range per $\mu\text{W}(\text{cm}^{-2}\text{nm})$	Scale Small $\mu\text{V}(\text{cm}^{-2}\text{nm})$	Scale Medium $\mu\text{V}(\text{cm}^{-2}\text{nm})$											
			Ed0306	2	305					3.5700E-14	4.0765E-06	1.1499E-03	-3.8203E-04	-5.3300E-04	-5.7100E-04 $\mu\text{W}(\text{cm}^{-2}\text{nm})$	
			Ed0310	8	313					1.7000E-10	1.7535E-06	5.0755E-03	1.7535E-03	2.2820E-04	2.4400E-04	1.5870E-03 $\mu\text{W}(\text{cm}^{-2}\text{nm})$
			Ed0320	8	320					2.5444E-10	2.5800E-06	7.6755E-03	2.4607E-03	-1.1000E-04	-1.3800E-04	1.4100E-04 $\mu\text{W}(\text{cm}^{-2}\text{nm})$
			Ed0340	10	340					1.0347E-10	1.0747E-06	3.7656E-03	2.0677E-03	1.8000E-04	2.0800E-04	1.6130E-03 $\mu\text{W}(\text{cm}^{-2}\text{nm})$
			Ed0350	12	350					7.2238E-11	7.3784E-06	2.1911E-03	7.4261E-04	-1.6900E-04	-1.5500E-04	4.1000E-05 $\mu\text{W}(\text{cm}^{-2}\text{nm})$
			Ed0364	13	365					3.0230E-10	3.0295E-06	8.0567E-03	3.0535E-03	-1.5600E-04	-1.5500E-04	-3.5100E-04 $\mu\text{W}(\text{cm}^{-2}\text{nm})$
			Broadband Channel	Address	Wavelength [nm]					Responsivity			Offset Small [pA]	Offset Medium [pA]	Offset Large [pA]	Measurement Units
										Range per $\mu\text{E}(\text{cm}^{-2}\text{s})$	Scale Small $\mu\text{V}(\text{cm}^{-2}\text{s})$	Scale Medium $\mu\text{V}(\text{cm}^{-2}\text{s})$				
Ed0PAR	18	400-700	1.6747E-05	1.7075E-02	5.0000E-02	1.8492E-05	-4.3100E-05	-5.1000E-05	1.7470E-03 $\mu\text{E}(\text{cm}^{-2}\text{s})$							
Auxiliary Channel	Address	Wavelength	Responsivity	Scale Small	Scale Medium	Scale Large	Offset Small	Offset Medium	Offset Large	Measurement Units						
Ed0Temp	22	C	1.0000E+00	1.0000E-02	1.0000E-02	1.0000E-02	0.0000E+00	0.0000E+00	0.0000E+00	$^{\circ}\text{C}$						
Ed0Min	27	C	1.0000E+00	-2.5000E-01	-2.5000E-01	-2.5000E-01	0.0000E+00	0.0000E+00	0.0000E+00	V						

* Biospherical Instruments Inc., 5540 Riley Street, San Diego, California 92119 USA. Contact support@biospherical.com for more information.

Calibration Data Do Not Destroy

Page 2 of 2

CTD Sensors

CTD Fish

SEA-BIRD ELECTRONICS, LLC.

13431 NE 20th Street, Bellevue, Washington, 9805-2010 USA

Phone: (425) 643-9886 Fax: (425) 643-0954 E-mail: seabird@seabird.com

SENSOR SERIAL NUMBER: 0377
CALIBRATION DATE: 22-Feb-11SBE3plus PRESSURE CALIBRATION DATA
10000 psia S/N 58949

DIGIQUARTZ COEFFICIENTS:

C1 = -4.840595e+004
 C2 = -2.017057e+003
 C3 = 1.464810e+002
 C4 = 1.990600e+002
 C5 = 0.000000e+000
 C6 = 2.438285e+001
 C7 = -2.150042e+001
 C8 = 3.353120e+006
 C9 = 2.132540e+009
 C0 = 0.000000e+000

AD590M, AD590B, SLOPE AND OFFSET:


AD590M = 1.1000e-002
 AD590B = -8.4004e+000
 Slope = 0.0090
 Offset = 0.0031482601

PRESSURE (PSIA)	INST OUTPUT(Hz)	INST INPUT(C)	INST OUTPUT(PSIA)	CORRECTED INST OUTPUT(PSIA)	RESIDUAL (PSIA)
14.553	33360.27	21.7	14.035	14.601	0.139
2014.815	34041.30	21.8	2014.183	2014.687	0.127
4014.666	34703.31	21.8	4014.553	4014.522	0.158
6014.392	35357.33	21.8	6014.311	6014.313	0.121
8014.375	35994.73	21.8	8014.405	8014.505	0.134
10014.454	36618.75	21.9	10014.420	10014.573	-0.075
12014.210	37234.66	21.9	12014.373	12014.178	-0.033
14014.342	37837.61	21.9	14014.359	14014.251	0.022
16014.706	38426.52	21.9	16014.272	16014.625	0.081
18014.734	39011.27	21.8	18014.358	18014.562	-0.172
14.553	33367.20	21.8	14.566	14.527	-0.031

Residual = corrected instrument pressure - reference pressure

Date, Avg Offset (psia):

22-Feb-11 0.00



Primary Temperature

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1080
CALIBRATION DATE: 11-Aug-09

SBE3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.79662156e-003
h = 6.73718098e-004
i = 2.68106485e-005
j = 2.22416936e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121360e-003
b = 6.00129222e-004
c = 1.50875983e-005
d = 2.22563642e-006
f0 = 5818.175


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	5818.175	-1.5001	0.00001
0.9999	6153.460	0.9999	-0.00002
4.4999	6645.971	4.4999	0.00003
7.9999	7166.094	7.9999	0.00000
11.4999	7714.567	11.4999	-0.00003
14.9999	8292.110	14.9999	0.00003
18.4999	8899.373	18.4999	-0.00001
21.9999	9537.035	21.9999	-0.00001
25.4999	10205.718	25.4999	-0.00001
28.9999	10906.037	28.9999	0.00003
32.4999	11638.541	32.4999	-0.00001

Temperature ITS-90 = $1/[g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]] - 273.15$ (°C)

Temperature IPTS-68 = $1/[a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]] - 273.15$ (°C)

Following the recommendation of JPOTS: T_{90} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature



Secondary Temperature

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1085
CALIBRATION DATE: 11-Aug-09

SBE3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.85864603e-003
h = 6.72265172e-004
i = 2.51449928e-005
j = 1.91993472e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121332e-003
b = 5.98717580e-004
c = 1.44426645e-005
d = 1.92130839e-006
f0 = 6441.591


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	6441.591	-1.5001	0.00003
0.9999	6813.677	0.9999	-0.00003
4.4999	7360.263	4.4999	-0.00003
7.9999	7937.536	7.9999	0.00001
11.4999	8546.301	11.4999	0.00003
14.9999	9187.360	15.0000	0.00009
18.4999	9861.419	18.4998	-0.00010
21.9999	10569.313	21.9998	-0.00005
25.4999	11311.724	25.5000	0.00007
28.9999	12089.278	28.9999	0.00000
32.4999	12902.681	32.4999	-0.00001

Temperature ITS-90 = $1/[g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]] - 273.15$ (°C)

Temperature IPTS-68 = $1/[a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]] - 273.15$ (°C)

Following the recommendation of JPOTS: T_{90} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature



Primary Conductivity

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3648
CALIBRATION DATE: 11-Sep-09

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHJ COEFFICIENTS

g = -1.01343555e+001
h = 1.55513334e+000
i = -2.53681791e-003
j = 2.82188355e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 4.64455701e-007
b = 1.54842242e+000
c = -1.01208117e+001
d = -8.11833808e-005
m = 6.4
CPcor = -9.5700e-008 (nominal)


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.55660	0.00000	0.00000
-0.9999	34.8745	2.80884	4.96666	2.80882	-0.00001
1.0001	34.8746	2.98048	5.07693	2.98050	0.00002
15.0001	34.8756	4.27813	5.84308	4.27811	-0.00002
18.5001	34.8742	4.62523	6.03143	4.62525	0.00002
29.0001	34.8711	5.71029	6.58511	5.71028	-0.00001
32.5002	34.8637	6.08334	6.76489	6.08335	0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / (10(1 + \delta t + \epsilon p))$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients



Secondary Conductivity

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0763
CALIBRATION DATE: 11-Aug-09

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHJ COEFFICIENTS

g = -4.21874180e+000
h = 4.46774452e-001
i = 1.16219632e-004
j = 1.35640404e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 5.98553293e-005
b = 4.46903159e-001
c = -4.21862812e+000
d = -7.65859402e-005
m = 3.6
CPcor = -9.5700e-008 (nominal)


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	3.07122	0.00000	0.00000
-1.0000	34.7667	2.80095	8.47470	2.80096	0.00001
1.0000	34.7671	2.97216	8.69616	2.97217	0.00001
15.0000	34.7687	4.26639	10.21431	4.26636	-0.00003
18.5000	34.7689	4.61276	10.58324	4.61274	-0.00002
29.0000	34.7669	5.69514	11.66002	5.69526	0.00012
32.5000	34.7623	6.06764	12.00765	6.06756	-0.00008

Conductivity = $(g + hf^2 + if^3 + jf^4) / [10(1 + \delta t + \epsilon p)]$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients



Primary Dissolved Oxygen

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMRFR: 0179
CALIBRATION DATE: 01-Mar-11p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.4685

Voffset = -0.4894

Tau20 = 1.35

A = -9.3328e-003

B = 3.6742e-004

C = -2.7405e-006

E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS

D1 = 1.92634e-4 H1 = -3.30000e-2

D2 = -4.64803e-2 H2 = 5.00000e+3


H3 = 1.45000e+3

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT (VOLTS)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.23	2.00	0.05	0.767	1.24	0.00
1.24	6.00	0.05	0.806	1.24	0.00
1.24	12.00	0.05	0.865	1.24	-0.00
1.25	20.00	0.05	0.935	1.25	-0.00
1.26	26.00	0.05	0.983	1.26	0.00
1.26	30.00	0.05	1.010	1.26	0.00
4.16	26.00	0.05	2.121	4.15	-0.00
4.16	30.00	0.05	2.207	4.16	-0.00
4.16	6.00	0.05	1.556	4.16	0.00
4.16	20.00	0.05	1.976	4.16	-0.00
4.16	2.00	0.05	1.424	4.16	-0.00
4.16	12.00	0.05	1.748	4.16	-0.00
6.59	30.00	0.05	3.213	6.59	-0.00
6.62	12.00	0.05	2.493	6.63	0.00
6.63	26.00	0.05	3.095	6.63	0.01
6.63	6.00	0.05	2.190	6.64	0.00
6.64	20.00	0.05	2.863	6.64	-0.00
6.66	2.00	0.05	1.985	6.66	-0.00

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * OxSol(T,S) * exp(E * P / K)

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen



Acquisition and Processing Information

Processing Specifics

Refer to the InstCoef.txt file along with the specific instrument calibration sheets, both located in the Cal/ directory of the data distribution, for information on how the RVDAS data was collected and processed.

Errors and Events

This section lists all significant events and known problems with acquisition during this cruise including instrument failures, data acquisition system failures, and other factors affecting this data set.

Date (Julian)	Time (GMT)	Event	Location
214	02:59	Started data collection	@68W
214	17:24	Changed out TSG	
219	16:40	Arrived Palmer Station	
224	13:57	Departed Palmer Station	
230	22:18	Shut down data collection	Entering Chilean EEZ

- TSG was changed out because the readings were very low (32PSU). Turns out those readings were correct, and were likely a result of the heavy snowfall melting off into the Straights of Magellan.
- Upon entering ice at approx 59° 30' South, seawater systems became frustratingly unreliable. Pumps frequently became airborne and lost pressure, and strainers were frequently clogged with krill.